# DR 4200, DR4300, DR 4500 & DR 4500A Circular Chart Recorders

Latch Assembly Kit 30756584-001 Lock Assembly Kit 30756584-002 *KIT INSTRUCTION 44-45-33-33B* 


### Summary

The assembly kit contains all the components you need to replace the latch assembly on the door of a DR 4200, DR4300, DR 4500 or DR 4500A recorder. Table 1 lists contents of the kit. Note that three housing retainers are supplied - one is for use with DR 4200 recorders; one is for older DR 4500 & DR 4500A recorders and one is for newer DR4500A & DR4300 Recorders. Use the appropriate procedure to install the kit. The procedures assume that the chart door is opened, the chart plate is swung out, and the power is removed.

Table 1 - Kit Contents


		LATCH 30756584-001	LOCK 30756584-002
Component	Part Number	Qty	Qty
Latch	30756100-001	1	0
Barrel	30756208-001	1	0
O-ring	30756216-001	1	0
Retaining Nut	30756209-001	1	0
Washer - Delrin	30756218-001	1	1
Latch Retainer	30756102-001	1	1
Housing Retainer – Older (DR 4500 & DR 4500A)	30756204-001	1	1
Housing Retainer – Newer (DR4500A & DR4300)	51205576-001	1	1
Housing Retainer (DR 4200)	30756101-001	1	1
Screw (10-32)	30674113-012	1	1
Cam	30756207-001	1	0
Screw (#10)	3000005-004	1	0
Washer	30004950-000	1	0
Lock Cylinder Assembly	30756047-001	0	1
Washer - S. Steel	30756218-002	1	1

### Removal Procedure



## Installation - Latch Assembly


### Installation Procedure - Latch Assembly



# Installation - Lock Assembly

Installation Procedure - Latch Assembly

# Installing Lock Assembly Kit NOTE: Lock assembly supplied may require some disassembly prior to performing installation procedure. 1. Put round washer (30756218-002) on lock cylinder (30756047-001). NOTE: If steel door use washer 30756218-001. 2. Insert lock cylinder through hole from front of door.



- 6. For proper installation, key should turn downward to lock door.
  - \* Component supplied as part of the lock assembly.

### Installing Latch on Chart Plate

- Remove rubber O-ring from back of captive screw in chart plate.
- Position retaining latch (30756102-001)
  in rectangular cutout of chart plate. Insert
  10-32 screw (30674113-012) through
  retaining latch and slip rubber O-ring over
  screw from back of chart plate.
- 3. Close chart plate and secure with captive screw.

NOTE: Always secure chart plate to case before engaging lock.

4. Close and lock door.

# Honeywell

**Industrial Automation and Control** 

Honeywell, Inc. 1100 Virginia Drive Fort Washington, Pennsylvania 19034 Helping You Control Your World